

CERTIFICATE OF CALIBRATION

Issued By Associated Calibration & Training Ltd

Date of Issue 24 April 2023

Certificate Number
ACT59381

Page 1 of 2 Pages



Associated Calibration & Training Ltd
Unit 12, North Staffs Business Park
Stoke-on-Trent
ST6 4BF
Tel: 01782 827018

Approved Signatory

T.P Rowley J. Taylor A.J Corbishley C. Whalley P.G Casey

Customer : Dave Jones (Oswestry)
4 Lime Grove, Oswestry
Shropshire SY11 2QE

Date Received : 24 April 2023

Instrument -	System ID :	ID17643	Job Number :	R59906-1
	Description :	Portable Appliance Tester		
	Manufacturer :	Seaward	Site :	
	Model Number :	PT100	Location :	
	Serial Number :	18K-0065	Last Certificate Number :	ACT47094
	Procedure Version :	49 3.00	Last Calibration Date :	07/04/2022

Environmental Conditions

Temperature :	21°C +/- 3°C	Mains Voltage :	240V +/- 10V
Relative Humidity :	50% +/- 15%	Mains Frequency :	50Hz +/- 1Hz

Comments

The Instrument was allowed to stabilise before calibration.

Traceability Information

<i>Instrument description</i>	<i>Serial number</i>	<i>Certificate number</i>	<i>Cal. Date</i>	<i>Cal. Period</i>
3200 Electrical Test Calibrator (HI)	M1405B16	UKAS47997	13/02/2023	52

Calibrated By : P.G Casey

Date of Calibration : 24 April 2023

This certificate provides traceability of measurement to recognised National Standards, and to the units of measurement realised at the National Physical Laboratory or other recognised National Standards laboratories.

Copyright of this certificate is owned by the issuing laboratory and may not be reproduced except with the prior written approval of the issuing laboratory. This certificate complies with the requirements of BS EN ISO 10012:2003.

CERTIFICATE OF CALIBRATION

Certificate Number
ACT59381

Page 2 of 2 Pages

Test Title	Tolerance	Applied Value	Reading	Pass/Fail
General Operation				
Display	---	---	Pass	
Controls	---	---	Pass	
Class One				
Pass 1				
Linearity	22.5m Ω	0.10 Ω	0.09 Ω	Pass
1M Ω	70k Ω	1.00M Ω	1.00M Ω	Pass
0.23mA	43uA	0.230mA	0.230mA	Pass
Pass 2				
Linearity	27.5m Ω	0.17 Ω	0.17 Ω	Pass
2M Ω	120k Ω	2.00M Ω	2.00M Ω	Pass
0.12mA	32uA	0.120mA	0.120mA	Pass
Pass 3				
Linearity	48m Ω	0.54 Ω	0.54 Ω	Pass
5M Ω	270k Ω	5.00M Ω	5.01M Ω	Pass
Class Two				
Pass 1				
2M Ω	120k Ω	2.00M Ω	2.00M Ω	Pass
0.12mA	32uA	0.120mA	0.120mA	Pass
Test 2				
10M Ω	520k Ω	10.00M Ω	10.01M Ω	Pass
Extension Lead Test				
Polarity Test	---	---	Pass	
End of results				

Uncertainties

Bond Current	0 to 30 Amps: 1.5% + 6counts.
Earth Bond Ohms	0.5% + 4mohms
Insulation Ohms	10kohms to 5Mohms: 0.1%, 5Mohms to 10Gohms: 1% +/- 1 Digit.
PAT Leakage	1.5% + 0.3mA